

AD000034

Documentation • Test report

Pressure documentation, DAkkS calibration for DW5x ≤400 bar, absolute pressure, traceable and accredited according to ISO 17025



- / 3 sets of measurements, each with 9 measurement points
- / Documentation in German and English

Global Recognition of the Calibration Certificate To ensure consistent product quality

A DAkkS calibration is a highly precise and internationally recognized comparison of measuring instruments with certified standards. It is performed exclusively by government-supervised and accredited laboratories in accordance with the DIN EN ISO/IEC 17025 standard. The procedure precisely documents the respective measurement uncertainty. The calibration certificate documents metrological traceability to national standards for the representation of units in accordance with international systems of units (SI).

The calibration certificate can only be ordered in conjunction with a device. The device is shipped together with the corresponding calibration certificate. The calibration certificate is bilingual (German/English).

Classification

ETIM 8

More

IPF Product Group	731 Special devices (miscellaneous)
packaging dimensions	
gross weight	
customs tariff number	90269000
WEEE number	40951076
Reach-compliant	Yes
RoHS-compliant	Yes

Important notes

- / the calibration certificate must be ordered together with a sensor. The corresponding sensor is delivered with the calibration certificate. The calibration certificate is bilingual (German/English).



Installation

Mounting / installation may only be carried out by a qualified electrician!



Disposal

WEEE number according to § 6 para. 3 ElektroG: 40951076

Safety warnings

- / Before initial operation, please make sure to follow all safety instructions that may be provided in the product information.
- / Never use these devices in applications where the safety of a person depends on their functionality.